

QUALITY CONTROL CERTIFICATE

KIT NAME DEAMIDATED GLIADIN-G - CHORUS

Code	86100		
Lot No.	987	2024-11	2023-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D8660	Deamidated Gliadin-G	987	2024-11	6X6
B8660	CALIBRATORE Deamidated Gliadin-G	987	2024-11	1
C8660	CONTROL SERUM Deamidated Gliadin-G	987	2024-11	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	481	≥ 100
Control Serum (AU/mL)	91,2	46,5-160,0

CALIBRATOR

2	5	5	9	8	7	6	1	2	0	2	8	3	8	4	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature MB

255987612028364800
DEAMIDATED GLIADIN-G
CALIBRATOR 0,175 mL IVD
LOT 987
 2024-11 2-8°C
DIESSE Monteriggioni Italy CE

CONTROL SERUM

2	5	5	9	8	7	6	1	9	3	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature MB

255987619340000001
DEAMIDATED GLIADIN-G
CONTROL+ 0,425 mL IVD
LOT 987
 2024-11 2-8°C
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents [Signature]

DATE 17/11/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.