

QUALITY CONTROL CERTIFICATE

KIT NAME dsDNA-G - CHORUS

Code	86032		
Lot No.	981	2024-10	2023-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D3281	dsDNA-G	981	2024-10	6X6
B3281	CALIBRATORE dsDNA-G	981	2024-10	1
C3281	CONTROL SERUM dsDNA-G	981	2024-10	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	650	≥ 100
Control Serum (IU/mL)	47,0	35,7-77,8

CALIBRATOR

2	5	1	9	8	1	6	6	2	4	3	3	4	5	3	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature IB

251981662433453600
dsDNA-G
CALIBRATOR 0.175 mL [IVD]
LOT 981
2024-10 2°C-8°C
DIESSE Monteriggioni Italy CE

CONTROL SERUM

2	5	1	9	8	1	6	1	6	2	9	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature IB

251981616290000001
dsDNA-G
CONTROL+ 0.425 mL [IVD]
LOT 981
2024-10 2°C-8°C
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents IB

DATE 23.11.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.