

**QUALITY CONTROL CERTIFICATE**

**KIT NAME** CHIKUNGUNYA IgM -CHORUS

Code	81134		
Lot No.	987	2024-11	2023-11

**REAGENTS**

Label	Description	Lot.	Exp.	NO
D0211	CHIKUNGUNYA IgM	987	2024-11	6x6
B0211	CALIBRATORE CHIKUNGUNYA IgM	987	2024-11	1
C0211	CONTROL CHIKUNGUNYA IgM	987	2024-11	1

**CONTROLS**

	Found results	Expected results
Calibrator (mOD)	668	≥ 100
Control (INDEX)	2,7	1,6-3,0

**CALIBRATOR**

9	1	0	9	8	7	6	0	0	0	0	4	3	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature RB



**CONTROL SERUM**

9	1	0	9	8	7	6	1	6	3	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature RB



**This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents Amor

DATE 23/11/2023

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.