

DIESE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME CHLAMYDOPHILA PNEUMONIAE IgG - CHORUS

Code	81250		
Lot No.	987	2024-11	2023-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D8590	Chlamydomphila pneumoniae IgG	987	2024-11	6X6
B8590	CALIBRATORE Chlamydomphila p. IgG	987	2024-11	1
C8590	CONTROL SERUM Chlamydomphila p. IgG	987	2024-11	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	450	≥ 100
Control Serum (AU/mL)	56,3	31,4-100,0
Negative Control (AU/mL)	<10,0	≤ 10,0

CALIBRATOR

2	2	2	9	8	7	6	7	1	6	3	0	2	9	4	5	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature RB

222987671630294500
CHLAMYDOPHILA
PNEUMONIAE IgG
CALIBRATOR 0,175 mL [IVD]
LOT 987
2024-11 2°C-8°C
DIESE Monteriggioni Italy CE 0123 B8590

CONTROL SERUM

2	2	2	9	8	7	6	2	4	4	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature RB

222987624450000001
CHLAMYDOPHILA
PNEUMONIAE IgG
CONTROL+ 0,425 mL [IVD]
LOT 987
2024-11 2°C-8°C
DIESE Monteriggioni Italy CE 0123 C8590

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents [Signature]

DATE 24/11/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.