



DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020

**QUALITY CONTROL CERTIFICATE**

**KIT NAME** CARDIOLIPIN-G - CHORUS

Code	86046		
Lot No.	981-A	2024-10	2023-11

**REAGENTS**

Label	Description	Lot.	Exp.	NO
D3311	CARDIOLIPIN-G	981	2024-10	6X6
B3311	CALIBRATORE CARDIOLIPIN-G	981-1	2024-11	1
C3311	CONTROL SERUM CARDIOLIPIN-G	981-1	2024-11	1

**CONTROLS**

	Found results	Expected results
Calibrator (mOD)	460	≥ 100
Control Serum (GPLU/mL)	39,1	28,2-72,6

**CALIBRATOR**

3	3	5	9	8	1	6	3	2	7	4	1	2	8	6	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *[Signature]*

335981632741286000  
Cardiolipin-G

**CALIBRATOR** 0.175 mL **IVD**

**LOT** 981-1

2024-11 2°C-8°C

**DIESSE** Monteriggioni Italy **CE**

B3311

**CONTROL SERUM**

3	3	5	9	8	1	6	2	5	4	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *[Signature]*

335981625470000001  
Cardiolipin-G

**CONTROL+** 0.425 mL **IVD**

**LOT** 981-1

2024-11 2°C-8°C

**DIESSE** Monteriggioni Italy **CE**

C3311

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *[Signature]*

DATE *20.11.2023*

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.