



QUALITY CONTROL CERTIFICATE

KIT NAME BRUCELLA IgM CHORUS

Code	81166		
Lot No.	987	2024-11	2023-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D4120	Brucella IgM	987	2024-11	6X6
B4120	CALIBRATORE Brucella IgM	987	2024-11	1
C4120	Control Serum Brucella IgM	987	2024-11	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	550	≥ 100
Control (Index)	3,7	2,6-4,8

CALIBRATOR

1	4	7	9	8	7	6	0	0	0	0	0	2	7	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *MB*

14798760000270000
Brucella IgM
CALIBRATOR 0.175 mL IVD
LOT 987
 2024-11 2°C-8°C
DIESSE Monteriggioni Italy CE
B4120

CONTROL SERUM

1	4	7	9	8	7	6	2	6	4	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *MB*

14798762648000001
Brucella IgM
CONTROL 0.425 mL IVD
LOT 987
 2024-11 2°C-8°C
DIESSE Monteriggioni Italy CE
C4120

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE *06.12.2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.