

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME RF-G - CHORUS

Code 86038

Lot No. 988

2024-11

2023-12

## REAGENTS

Label	Description	Lot.	Exp.	NO
D7411	RF-G	988	2024-11	6X6
B7411	CALIBRATORE RF-G	988	2024-11	1
C7411	CONTROL SERUM RF-G	988	2024-11	1

## CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	453	$\geq 100$
Control Serum (AU/mL)	45,9	29,0-73,1


## CALIBRATOR

1	6	1	9	8	8	6	3	2	5	3	8	4	3	5	2	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature



## CONTROL SERUM

1	6	1	9	8	8	6	1	6	3	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 12.12.2023

## WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.