



DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME HSV 2 SCREEN - CHORUS

Code	81019		
Lot No.	987	2025-02	2023-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D2201	HSV 2 SCREEN	987	2025-02	6x6
B2201	CALIBRATORE HSV 2 SCREEN	987	2025-04	1
C2201	CONTROL SERUM HSV 2 SCREEN	987	2025-06	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	405	≥ 100
Control Serum (index)	2,5	1,3-8,0
Negative Control (index)	0,6	< 0,7

CALIBRATOR

0	3	4	9	8	7	7	0	0	0	0	0	5	3	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

034987700000530000
HSV 2 Screen
CALIBRATOR 0.3 mL IVD
LOT 987
2025-04
DIESSE Monteriggioni Italy CE

CONTROL SERUM

0	3	4	9	8	7	7	1	3	8	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

034987713800000001
HSV 2 Screen
CONTROL 0.5 mL IVD
LOT 987
2025-06
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents...

DATE 13/12/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.