



QUALITY CONTROL CERTIFICATE

KIT NAME LKM-1 CHORUS

Code	86080		
Lot No.	996	2024-12	2023-12

REAGENTS

Label	Description	Lot.	Exp.	NO
D3641	LKM-1	996	2024-12	6X6
B3641	CALIBRATORE LKM-1	996	2024-12	1
C3641	CONTROL SERUM LKM-1	996	2024-12	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	500	≥ 100
Control Serum (AU/mL)	72,6	46,1-100,0

CALIBRATOR

7	0	3	9	9	6	6	4	1	9	3	3	4	0	5	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *FB*

703996641933405900  
LKM-1  
CALIBRATOR 0.175 mL IVD  
LOT 996  
 2024-12 2°C 8°C  
DIESSA Monteriggioni Italy CE

CONTROL SERUM

7	0	3	9	9	6	6	1	8	3	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *FB*

703996618330000001  
LKM-1  
CONTROL + 0.425 mL IVD  
LOT 996  
 2024-12 2°C 8°C  
DIESSA Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents: *FB*

DATE 14/12/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!  
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.