

DIESE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME RUBELLA IgM - CHORUS

Code	81031			
Lot No.	988	2024-11	2023-11	

REAGENTS

Label	Description	Lot.	Exp.	NO
D2241	RUBELLA IgM	988	2024-11	6X6
B2241	CALIBRATORE RUBELLA IgM	988	2024-11	1
C2241	CONTROL SERUM	988	2024-11	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	545	≥ 100
Control Serum (INDEX)	2,7	1,8-3,3
Negative Control (INDEX)	0,0	< 0,7

CALIBRATOR

1	3	5	9	8	8	6	0	0	0	0	0	4	0	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature MB

13598860000400000
Rubella IgM
CALIBRATOR 0.175 mL IVD
LOT 988
2024-11
2°C - 8°C
DIESE Monteriggioni Italy CE 0123 B2241

CONTROL SERUM

1	3	5	9	8	8	6	1	8	3	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature MB

135988618330000001
Rubella IgM
CONTROL+ 0.425 mL IVD
LOT 988
2024-11
2°C - 8°C
DIESE Monteriggioni Italy CE 0123 C2241

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... [Signature]

DATE 18/12/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.