

QUALITY CONTROL CERTIFICATE

KIT NAME ANA-Screen - CHORUS

Code	86014		
Lot No.	976	2024-12	2023-12

REAGENTS

Label	Description	Lot.	Exp.	NO
D3141	ANA-Screen	976	2024-12	6X6
B3141	CALIBRATORE ANA-Screen	976	2024-12	1
C3141	CONTROL SERUM ANA-Screen	976	2024-12	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	368	≥ 100
Control Serum (Ndx)	2,6	2,0-3,6

CALIBRATOR

1	1	5	9	7	6	7	0	0	0	0	0	3	6	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

11597670000360000
ANA-Screen
CALIBRATOR 0.175 mL IVD
LOT 976
2024-12 2°C-8°C
DIESSE Monteriggioni Italy CE
B3141

CONTROL SERUM

1	1	5	9	7	6	7	2	0	3	6	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

115976720360000001
ANA-Screen
CONTROL+ 0.425 mL IVD
LOT 976
2024-12 2°C-8°C
DIESSE Monteriggioni Italy CE
C3141

*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 20/12/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.