

QUALITY CONTROL CERTIFICATE

KIT NAME RESPIRATORY SYNCYTIAL VIRUS IgG - CHORUS

Code	81036		
Lot No.	996	2024-12	2023-12

REAGENTS

Label	Description	Lot.	Exp.	NO
D7661	Respiratory Syncytial Virus IgG	996	2024-12	6X6
B7661	CALIBRATORE Respiratory Syncytial Virus IgG	996	2024-12	1
C7661	CONTROL SERUM Respiratory Syncytial Virus IgG	996	2024-12	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	480	≥ 100
Control Serum (Ndx)	1,9	1,5-2,8
Negative Control (Ndx)	0,0	≤ 0,7

CALIBRATOR

1	5	9	9	9	6	6	0	0	0	0	0	4	7	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *Deus*



CONTROL SERUM

1	5	9	9	9	6	6	1	5	2	8	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *Deus*



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Deus*

DATE 22-12-2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.