

**QUALITY CONTROL CERTIFICATE**

**KIT NAME** CALPROTECTIN - CHORUS

Code	86600		
Lot No.	976	2024-09	2023-09

**REAGENTS**

Label	Description	Lot.	Exp.	NO
D9500	CALPROTECTIN	976	2024-09	2 X 6
B9500	CALIBRATORE CALPROTECTIN	976	2025-01	1
C9500	CONTROL CALPROTECTIN	976	2025-01	1
G9502	SAMPLE DILUENT	976	2025-09	1


**CONTROLS**

	Found results	Expected results
Calibrator (mOD)	420	≥ 100
Control (mg/kg)	162,6	87,5-250,0

**CALIBRATOR**

9	0	0	9	7	6	6	2	1	8	2	2	4	5	2	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature FB



**CONTROL SERUM**

9	0	0	9	7	6	6	1	6	2	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature FB



**This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... [Signature]

DATE 05/01/2024

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.