

DIESSA DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME TETANUS IgG - CHORUS

Code	81270		
Lot No.	988	2024-11	2023-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D9660	Tetanus IgG	988	2024-11	6X6
B9660	CALIBRATORE Tetanus IgG	988	2024-11	1
C9660	CONTROL SERUM Tetanus IgG	988	2024-11	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	583	≥ 100
Control (mIU/mL)	664,2	431,8-1000,0

CALIBRATOR

9	0	2	9	8	8	6	3	2	1	3	3	4	0	5	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature MB

902988632133405000
Tetanus IgG
CALIBRATOR 0.175 mL IVD
LOT 988
2024-11
DIESSA Monteriggioni Italy CE

CONTROL SERUM

9	0	2	9	8	8	6	1	8	3	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature MB

902988618330000001
Tetanus IgG
CONTROL+ 0.425 mL IVD
LOT 988
2024-11
DIESSA Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents [Signature]

DATE 09/01/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.