



QUALITY CONTROL CERTIFICATE

KIT NAME Epstein-Barr VCA IgG - CHORUS

Code	81055		
Lot No.	989	2025-02	2023-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D2341	Epstein-Barr VCA IgG	989	2025-02	6X6
B2342	CALIBRATORE Epstein-Barr VCA IgG	989	2025-02	1
C2341	CONTROL SERUM Epstein-Barr VCA IgG	989	2025-02	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	336	≥ 100
Control Serum (INDEX)	4,8	3,3-6,2
Negative Control (INDEX)	0,1	<0,7

CALIBRATOR

0	0	5	9	8	9	7	0	0	0	0	0	2	1	0	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *88*



CONTROL SERUM

0	0	5	9	8	9	7	3	3	6	2	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *88*



*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *88*

DATE *10.01.2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.