

QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS ECHINOCOCCUS IgG

Code	81284		
Lot No.	001	2025-01	2024-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D0010	Echinococcus IgG	001	2025-01	2X6
B0010	CALIBRATOR Echinococcus IgG	001	2025-01	1
C0010	CONTROL SERUM Echinococcus IgG	001	2025-01	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	860	≥ 100
Control Serum (Index)	1,6	1,3-2,3

CALIBRATOR

0	3	5	0	0	1	6	0	0	0	0	0	5	6	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

03500160000560000
Echinococcus IgG
CALIBRATOR 0.175 mL IVD
LOT 001
2025-01 2°C-8°C
DIESSA Monteriggioni Italy CE

CONTROL SERUM

0	3	5	0	0	1	6	1	3	2	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

03500161323000001
Echinococcus IgG
CONTROL+ 0.425 mL IVD
LOT 001
2025-01 2°C-8°C
DIESSA Monteriggioni Italy CE

*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents Abus

DATE 22/01/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.