



QUALITY CONTROL CERTIFICATE

KIT NAME SS-A - CHORUS

Code	86018		
Lot No.	001	2025-04	2024-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D3181	SS-A	001	2025-04	6X6
B3181	CALIBRATORE SS-A	001	2025-04	1
C3181	CONTROL SERUM SS-A	001	2025-04	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	416	≥ 100
Control Serum (AU/mL)	54,4	28,9-94,3

CALIBRATOR

1	9	3	0	0	1	7	1	2	4	3	6	3	1	4	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*



CONTROL SERUM

1	9	3	0	0	1	7	2	3	4	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Abun*

DATE *19/10/2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.