

DIESE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME CARDIOLIPIN-G - CHORUS

Code	86046		
Lot No.	002	2025-01	2024-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D3311	CARDIOLIPIN-G	002	2025-01	6X6
B3311	CALIBRATORE CARDIOLIPIN-G	002	2025-01	1
C3311	CONTROL SERUM CARDIOLIPIN-G	002	2025-01	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	500	≥ 100
Control Serum (GPLU/mL)	27,5	21,3-40,0

CALIBRATOR

3	3	5	0	0	2	6	5	2	9	4	4	3	8	6	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

335002652944386400
Cardiolipin-G
CALIBRATOR 0.175 mL IVD
LOT 002
2025-01 2°C-8°C
DIESE Monteriggioni Italy CE

CONTROL SERUM

3	3	5	0	0	2	6	1	8	3	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

335002618340000001
Cardiolipin-G
CONTROL+ 0.425 mL IVD
LOT 002
2025-01 2°C-8°C
DIESE Monteriggioni Italy CE

**This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents [Signature]

DATE 16/01/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.