



QUALITY CONTROL CERTIFICATE

KIT NAME HELICOBACTER PYLORI IgA-CHORUS

Code	81062		
Lot No.	990	2024-11	2023-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D2401	Helicobacter Pylori IgA	990	2024-11	6X6
B2401	CALIBRATORE Helicobacter Pylori IgA	990	2024-11	1
C2401	CONTROL SERUM Helicobacter Pylori IgA	990	2024-11	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	370	≥ 100
Control Serum (INDEX)	3,4	2,3-4,2
Negative Control (INDEX)	0,0	<0,7

CALIBRATOR

0	2	7	9	9	0	6	0	0	0	0	0	3	1	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*

02799060000310000
Helicobacter pylori IgA
CALIBRATOR 0.178 mL IVD
LOT 990
 2024-11 2°C-8°C
DIESSE Monteriggioni Italy CE
B2401

CONTROL SERUM

0	2	7	9	9	0	6	2	3	4	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*

02799062342000001
Helicobacter pylori IgA
CONTROL+ 0.428 mL IVD
LOT 990
 2024-11 2°C-8°C
DIESSE Monteriggioni Italy CE
C2401

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *FB*

DATE *23/01/2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.