



DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

**QUALITY CONTROL CERTIFICATE**

**KIT NAME** DEAMIDATED GLIADIN-A - CHORUS

Code	86098		
Lot No.	989	2024-11	2023-11

**REAGENTS**

Label	Description	Lot.	Exp.	NO
D8650	Deamidated Gliadin-A	989	2024-11	6X6
B8650	CALIBRATORE Deamidated Gliadin-A	989	2024-11	1
C8650	CONTROL SERUM Deamidated Gliadin-A	989	2024-11	1

**CONTROLS**

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	541	≥ 100
Control Serum (AU/mL)	60,5	38,3-92,9

**CALIBRATOR**

2	5	4	9	8	9	6	3	2	5	3	9	3	5	5	5	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature HB

254989632539355500  
DEAMIDATED GLIADIN-A  
CALIBRATOR 0.175 mL [I]VD  
LOT 989  
 2024-11 2023-11  
DIESSE Monteriggioni Italy CE

**CONTROL SERUM**

2	5	4	9	8	9	6	2	0	3	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature HB

254989620370000001  
DEAMIDATED GLIADIN-A  
CONTROL 0.425 mL [I]VD  
LOT 989  
 2024-11 2023-11  
DIESSE Monteriggioni Italy CE

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents HB

DATE 25/01/2024

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.