

QUALITY CONTROL CERTIFICATE

KIT NAME CALPROTECTIN - CHORUS

Code	86600			
Lot No.	001	2025-01	2024-01	

REAGENTS

Label	Description	Lot.	Exp.	NO
D9500	CALPROTECTIN	001	2025-01	2 X 6
B9500	CALIBRATORE CALPROTECTIN	001	2025-01	1
C9500	CONTROL CALPROTECTIN	001	2025-01	1
G9502	SAMPLE DILUENT	001	2026-01	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	300	≥ 100
Control (mg/kg)	138,3	86,1-250,0

CALIBRATOR

9	0	0	0	0	1	6	3	2	8	3	3	3	0	4	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature FB

90000163283304300  
Calprotectin  
CALIBRATOR 0,300 mL IVD  
LOT 001  
 2025-01 2°C-8°C  
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

9	0	0	0	0	1	6	2	3	4	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature FB

90000162343000001  
Calprotectin  
CONTROL 0,600 mL IVD  
LOT 001  
 2025-01 2°C-8°C  
DIESSSE Monteriggioni Italy CE

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... FB

DATE 26/01/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.