

**QUALITY CONTROL CERTIFICATE**

**KIT NAME 88000-W6 MUGWORT**

Code	88000-W6		
Lot No.	996	2024-12	2024-01

**REAGENTS**

Label	Description	Lot.	Exp.	NO
D3481	88000-W6 MUGWORT	996	2024-12	2X6
B3480	CALIBRATORE 88000-W6 MUGWORT	996	2025-02	1
C3480	CONTROL SERUM 88000-W6 MUGWORT	996	2025-02	1

**CONTROLS**

	Found results	Expected results
Calibrator (mOD)	310	≥ 100
Control Serum (KU/L)	9,9	5,8-14,3

**CALIBRATOR**

6	2	3	9	9	6	6	5	2	1	6	0	2	6	8	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature FB

623996652160268100  
Mugwort  
CALIBRATOR 0.250 mL IVD  
LOT 996  
2025-02  
2°C-8°C  
DIESSA Monteriggioni Italy CE

**CONTROL SERUM**

6	2	3	9	9	6	6	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature FB

623996627500000001  
Mugwort  
CONTROL+ 0.450 mL IVD  
LOT 996  
2025-02  
2°C-8°C  
DIESSA Monteriggioni Italy CE

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. Agno

DATE 26.02.24

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.