

DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME FT4 - CHORUS

Code	86520		
Lot No.	001	2025-01	2024-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D3320	ft4	001	2025-01	6X6
B3320	CALIBRATOR FT4	001	2025-01	1
C3320	CONTROL SERUM FT4	001	2025-01	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	360	≥ 100
Control Serum (pmol/L)	27,2	16,7-34,8

CALIBRATOR

9	6	8	0	0	1	6	3	2	9	3	9	3	6	4	5	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *MB*

968001632939364500
FT4
CALIBRATOR 0,250 mL IVD
LOT 001
2025-01 2°C-8°C
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

9	6	8	0	0	1	6	2	0	3	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *MB*

968001620360000001
FT4
CONTROL+ 0,500 mL IVD
LOT 001
2025-01 2°C-8°C
DIESSSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *MB*

DATE *26 d. 2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.