



DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME BORDETELLA PERTUSSIS TOXIN IgG - CHORUS

Code	81272		
Lot No.	001	2025-01	2024-01

REAGENTS

Label	Description	Lot.	Exp	No
D2100	BORDETELLA PERTUSSIS TOXIN IgG	001	2025-01	6x6
B2100	CALIBRATORE BORDETELLA P.TOXIN IgG	001	2025-06	1
C2100	CONTROL SERUM BORDETELLA P.TOXIN IgG	001	2025-06	1

CONTROLS

	Found results	Expected results
CALIBRATOR (mOD)	429	>100
CONTROL SERUM (IU/ml)	196,6	115,4-320,0

CALIBRATOR

9	6	4	0	0	1	6	1	2	0	3	3	3	3	4	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *SS*

964001612033334300
Bordetella pertussis
Toxin IgG
CALIBRATOR 0.450 mL [IVD]
LOT 001
2025-06 2°C-8°C
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

9	6	4	0	0	1	6	2	2	4	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *SS*

964001622410000001
Bordetella pertussis
Toxin IgG
CONTROL+ 0.900 mL [IVD]
LOT 001
2025-06 2°C-8°C
DIESSSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE *30.01.2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.