



QUALITY CONTROL CERTIFICATE

KIT NAME BORDETELLA PERTUSSIS TOXIN IgA - CHORUS

Code 81274
Lot No. 001 2025-01 2024-01

REAGENTS

Label	Description	Lot.	Exp	No
D3440	BORDETELLA PERTUSSIS TOXIN IgA	001	2025-01	6X6
B3440	CALIBRATORE BORDETELLA P.TOXIN IgA	001	2025-06	1
C3440	CONTROL SERUM BORDETELLA P.TOXIN IgA	001	2025-06	1

CONTROLS

	Found results	Expected results
CALIBRATOR (mOD)	500	>100
CONTROL SERUM (IU/ml)	26,7	16,9-33,3

CALIBRATOR

9 6 5 0 0 1 6 1 1 8 3 0 6 5 4 2 0 0

Label checked with barcode reader YES Operator Signature

965001611830654200
Bordetella pertussis
Toxin IgA
CALIBRATOR 0.450 mL IVD
LOT 001
2025-06 2°C-8°C
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

9 6 5 0 0 1 6 1 1 2 0 0 0 0 0 0 0 1

Label checked with barcode reader YES Operator Signature

965001611200000001
Bordetella pertussis
Toxin IgA
CONTROL 0.900 mL IVD
LOT 001
2025-06 2°C-8°C
DIESSSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 06/02/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.