

DIESSA DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME **TOXOPLASMA IgG AVIDITY - CHORUS**

Code	81098		
Lot No.	001	🕒 2025-01	📄 2024-01

REAGENTS

Label	Description	Lot	Exp.	NO
D2531	TOXOPLASMA IgG AVIDITY	001	2025-01	2x6
C2531	CONTROL SERUM HIGH TOXOPLASMA IgG AVIDITY	001	2025-01	1
G2531	SAMPLE DILUENT	001	2025-04	1

CONTROLS

	Found results	Expected results
Control Serum (%)	72,2%	> 50 ≤ 90 %

CALIBRATOR

--	--	--	--	--	--	--	--	--	--	--	--	--	--


Label checked with barcode reader YES Operator Signature

STICK LABEL

CONTROL SERUM

0	8	3	0	0	1	6	5	0	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*




083001650900000001
Toxoplasma IgG Avidity

CONTROL HIGH 0.3 mL LVD

LOT 001

🕒 2025-01 2°C

DIESSA Monteriggioni Italy  8123

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *[Signature]*

DATE *19/01/2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.