

**QUALITY CONTROL CERTIFICATE**

**KIT NAME** MUMPS IgM CHORUS

Code	81076		
Lot No.	001	2025-01	2024-01

**REAGENTS**

Label	Description	Lot.	Exp.	NO
D2461	MUMPS IgM	001	2025-01	6x6
B2461	CALIBRATORE MUMPS IgM	001	2025-01	1
C2461	CONTROL SERUM MUMPS IgM	001	2025-01	1


**CONTROLS**

	Found results	Expected results
Calibrator (mOD)	430	≥ 100
Control Serum (Index)	2,0	1,4-2,7
Negative Control (Index)	0,1	< 0,7

**CALIBRATOR**

1	3	2	0	0	1	6	0	0	0	0	0	4	9	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *Spalini*



**CONTROL SERUM**

1	3	2	0	0	1	6	1	4	2	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *Spalini*



**This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Spalini*

DATE 07.02.2024

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.