



QUALITY CONTROL CERTIFICATE

KIT NAME BORDETELLA PERTUSSIS TOXIN IgG - CHORUS

Code	81272		
Lot No.	008	2025-02	2024-02

REAGENTS

Label	Description	Lot.	Exp	No
D2100	BORDETELLA PERTUSSIS TOXIN IgG	008	2025-02	6x6
B2100	CALIBRATORE BORDETELLA P.TOXIN IgG	008	2025-06	1
C2100	CONTROL SERUM BORDETELLA P.TOXIN IgG	008	2025-06	1

CONTROLS

	Found results	Expected results
CALIBRATOR (mOD)	520	>100
CONTROL SERUM (IU/ml)	195,5	133,3-320,0

CALIBRATOR

9	6	4	0	0	8	6	1	1	7	3	2	3	3	4	2	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature *88*



CONTROL SERUM

9	6	4	0	0	8	6	2	2	4	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *88*



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 09.07.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.