



DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME a-TPO - CHORUS

Code	86074		
Lot No.	001	2025-01	2024-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D3591	a-TPO	001	2025-01	6X6
B3591	CALIBRATORE a-TPO	001	2025-01	1
C3591	CONTROL SERUM a-TPO	001	2025-01	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	278	≥ 100
Control Serum (IU/mL)	234,4	221,4-478,6

CALIBRATOR

3 0 1 0 0 1 6 5 2 6 5 4 2 6 7 3 0 0

Label checked with barcode reader YES Operator Signature FB

301001652654267300
a-TPO
CALIBRATOR 0.175 mL IVD
LOT 001
2025-01 2°C-8°C
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

3 0 1 0 0 1 6 2 8 5 2 0 0 0 0 0 0 1

Label checked with barcode reader YES Operator Signature FB

301001628520000001
a-TPO
CONTROL+ 0.425 mL IVD
LOT 001
2025-01 2°C-8°C
DIESSSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents Joanele

DATE 13/02/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.