



QUALITY CONTROL CERTIFICATE

KIT NAME POLIOVIRUS IgG - CHORUS

Code	81264/12		
Lot No.	987	2024-11	2024-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D9630	Poliovirus IgG	987	2024-11	2X6
B9630	CALIBRATORE Poliovirus IgG	987	2024-12	1
C9630	CONTROL SERUM Poliovirus IgG	987	2024-12	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	473	≥ 100
Control (Index)	1,4	1,1-2,1

CALIBRATOR

9	0	1	9	8	7	6	0	0	0	0	0	6	2	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature SS

901987600000620000  
Poliovirus IgG  
CALIBRATOR 0,175 mL [IVD]  
LOT 987  
2024-12 2°C-8°C  
DIESSE Monteriggioni Italy CE

CONTROL SERUM

9	0	1	9	8	7	6	1	1	2	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature SS

901987611210000001  
Poliovirus IgG  
CONTROL+ 0,425 mL [IVD]  
LOT 987  
2024-12 2°C-8°C  
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... SS

DATE 11.02.24

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.