



QUALITY CONTROL CERTIFICATE

KIT NAME 88000-F49 APPLE

Code	88000-F49		
Lot No.	002	2025-01	2024-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D1061	88000-F49 APPLE	002	2025-01	2X6
B1060	CALIBRATORE 88000-F49 APPLE	002	2025-02	1
C1060	CONTROL SERUM 88000-F49 APPLE	002	2025-02	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	284	≥ 100
Control Serum (KU/L)	8,2	5,8-14,3

CALIBRATOR

6	6	0	0	0	2	6	5	2	1	6	0	2	6	8	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

660002652160268100
Apple
CALIBRATOR 0.280 mL [IVD]
LOT 002
 2025-02 2°C B1060
DIESSE Monteriggioni Italy CE

CONTROL SERUM

6	6	0	0	0	2	6	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

660002627500000001
Apple
CONTROL SERUM 0.480 mL [IVD]
LOT 002
 2025-02 2°C C1060
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 15.02.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.