



QUALITY CONTROL CERTIFICATE

KIT NAME TG - CHORUS

Code	86076		
Lot No.	002	2025-01	2024-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D3613	TG	002	2025-01	6x6
B3613	CALIBRATORE TG	002	2025-01	1
C3613	CONTROL SERUM TG	002	2025-01	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	437	≥ 100
Control Serum (ng/mL)	15,7	13,6-30,0

CALIBRATOR

4	2	0	0	0	2	6	5	2	7	4	5	4	0	6	5	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature FB



CONTROL SERUM

4	2	0	0	0	2	6	1	8	3	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB



**This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. FB

DATE 16/02/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.