

DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME DEAMIDATED GLIADIN-A - CHORUS

Code	86098		
Lot No.	007	2025-02	2024-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D8650	Deamidated Gliadin-A	007	2025-02	6X6
B8650	CALIBRATORE Deamidated Gliadin-A	007	2025-02	1
C8650	CONTROL SERUM Deamidated Gliadin-A	007	2025-02	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	334	≥ 100
Control Serum (AU/mL)	85,6	60,7-275,0

CALIBRATOR

2	5	4	0	0	7	6	3	2	5	3	9	2	5	5	5	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *[Signature]*

254007632539255500
DEAMIDATED GLIADIN-A
CALIBRATOR 0.175 mL [IVD]
LOT 007
2025-02 2°C-8°C
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

2	5	4	0	0	7	6	2	8	5	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *[Signature]*

254007628530000001
DEAMIDATED GLIADIN-A
CONTROL 0.425 mL [IVD]
LOT 007
2025-02 2°C-8°C
DIESSSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *[Signature]*

DATE 19.02.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.