

DIESSÉ DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME dsDNA-G - CHORUS

Code	86032		
Lot No.	007	2025-02	2024-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D3281	dsDNA-G	007	2025-02	6X6
B3281	CALIBRATORE dsDNA-G	007	2025-02	1
C3281	CONTROL SERUM dsDNA-G	007	2025-02	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	649	≥ 100
Control Serum (IU/mL)	43,1	31,3-54,2

CALIBRATOR

2	5	1	0	0	7	6	6	2	2	3	4	5	6	3	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*

251007662234563700
dsDNA-G
CALIBRATOR 0.175 mL [IVD]
LOT 007
2025-02
DIESSÉ Monteriggioni Italy
CE

CONTROL SERUM

2	5	1	0	0	7	6	1	3	2	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*

251007613230000001
dsDNA-G
CONTROL 0.425 mL [IVD]
LOT 007
2025-02
DIESSÉ Monteriggioni Italy
CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE *11.01.2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.