

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME ENA-6 S - CHORUS

|         |       |         |         |
|---------|-------|---------|---------|
| Code    | 86012 |         |         |
| Lot No. | 003   | 2025-01 | 2024-01 |

REAGENTS

| Label | Description           | Lot. | Exp.    | NO  |
|-------|-----------------------|------|---------|-----|
| D3131 | ENA-6 S               | 003  | 2025-01 | 6X6 |
| B3131 | CALIBRATORE ENA-6 S   | 003  | 2025-01 | 1   |
| C3131 | CONTROL SERUM ENA-S 6 | 003  | 2025-01 | 1   |

CONTROLS

|  | Found results | Expected results |
|--|---------------|------------------|
|--|---------------|------------------|

|                     |     |         |
|---------------------|-----|---------|
| Calibrator (mOD)    | 410 | ≥ 100   |
| Control Serum (ndx) | 2,5 | 1,6-3,0 |

CALIBRATOR

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 1 | 0 | 2 | 0 | 0 | 3 | 6 | 0 | 0 | 0 | 0 | 0 | 4 | 3 | 0 | 0 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES  Operator Signature

10200360000430000  
ENA-6 S  
CALIBRATOR 0.175 mL IVD  
LOT 003  
 2025-01 2°C

CONTROL SERUM

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 1 | 0 | 2 | 0 | 0 | 3 | 6 | 1 | 6 | 3 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES  Operator Signature

102003616300000001  
ENA-6 S  
CONTROL+ 0.425 mL IVD  
LOT 003  
 2025-01 2°C

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents:

DATE 20.07.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.