



DIESSSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME dsDNA-G - CHORUS

Code	86032		
Lot No.	008	2025-02	2024-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D3281	dsDNA-G	008	2025-02	6X6
B3281	CALIBRATORE dsDNA-G	008	2025-02	1
C3281	CONTROL SERUM dsDNA-G	008	2025-02	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	679	≥ 100
Control Serum (IU/mL)	38,5	30,0-46,7

CALIBRATOR

2	5	1	0	0	8	6	5	2	5	3	5	5	6	3	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES  Operator Signature *SS*



CONTROL SERUM

2	5	1	0	0	8	6	1	3	2	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *SS*



*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*  
DATE *22.02.2024*

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**  
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.