



QUALITY CONTROL CERTIFICATE

KIT NAME PR3 - CHORUS

| | | | |
|---------|-------|---------|---------|
| Code | 86066 | | |
| Lot No. | 007 | 2025-02 | 2024-02 |

REAGENTS

| Label | Description | Lot | Exp. | NO |
|-------|-------------------|-----|---------|-----|
| D3512 | PR3 | 007 | 2025-02 | 6X6 |
| B3512 | CALIBRATORE PR3 | 007 | 2025-02 | 1 |
| C3512 | CONTROL SERUM PR3 | 007 | 2025-02 | 1 |

CONTROLS

| | Found results | Expected results |
|--|---------------|------------------|
|--|---------------|------------------|

| | | |
|-----------------------|------|-----------|
| Calibrator (mOD) | 423 | ≥ 100 |
| Control Serum (AU/mL) | 59,2 | 47,5-95,0 |

CALIBRATOR

| | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 2 | 5 | 2 | 0 | 0 | 7 | 6 | 7 | 1 | 5 | 2 | 7 | 3 | 2 | 4 | 3 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES Operator Signature HB

252007671527324300
PR3
CALIBRATOR 0.175 mL IVD
LOT 007
 2025-02 2°C - 8°C
DIESSE Monteriggioni Italy CE
B3512

CONTROL SERUM

| | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 2 | 5 | 2 | 0 | 0 | 7 | 6 | 2 | 2 | 4 | 1 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES Operator Signature HB

252007622410000001
PR3
CONTROL+ 0.425 mL IVD
LOT 007
 2025-02 2°C - 8°C
DIESSE Monteriggioni Italy CE
C3512

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents [Signature]
DATE 22/02/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.