



QUALITY CONTROL CERTIFICATE

KIT NAME West Nile IgG -CHORUS

Code	81174		
Lot No.	007	2025-02	2024-02

REAGENTS

Label	Description	Lot	Exp	No
D4080	West Nile IgG	007	2025-02	6X6
B4080	Calibratore West Nile IgG	007	2025-02	1
C4080	Control West Nile IgG	007	2025-02	1

CONTROLS

Found results	Expected results
---------------	------------------

Calibrator (mOD)	4,70	≥ 100
Control (Index)	2,7	1,7-3,2

CALIBRATOR

9	8	8	0	0	7	6	0	0	0	0	0	4	1	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature MB

98800760000410000
West Nile Virus IgG
CALIBRATOR 0.175 mL IVD
LOT 007
2025-02 2°C-8°C
DIESSE Monteriggioni Italy CE

CONTROL SERUM

9	8	8	0	0	7	6	1	7	3	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature MB

988007617320000001
West Nile Virus IgG
CONTROL+ 0.425 mL IVD
LOT 007
2025-02 2°C-8°C
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents [Signature]

DATE 29.02.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.