

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME LEGIONELLA PNEUMOPHILA 1-6 IgG -CHORUS

Code 81090
Lot No. 007 2024-11 2024-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D7381	LEGIONELLA PNEUMOPHILA 1-6 IgG	007	2024-11	6X6
B7381	CALIBRATORE L. PNEUMOPHILA 1-6 IgG	007	2024-11	1
C7381	CONTROL SERUM L. PNEUMOPHILA 1-6 IgG	007	2024-11	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	370	≥ 100
Control Serum (INDEX)	3,0	2,1-4,0
Negative Control (INDEX)	0,0	<0,7

CALIBRATOR

0 6 3 0 0 7 5 0 0 0 0 0 3 3 0 0 0 0

Label checked with barcode reader YES Operator Signature FB

06300750000330000
Legionella pneumophila 1-6 IgG
CALIBRATOR 0.175 mL IVD
LOT 007
2024-11 2024-11 2°C-8°C
DIESSE Monteriggioni Italy CE

CONTROL SERUM

0 6 3 0 0 7 5 2 1 4 0 0 0 0 0 0 0 1

Label checked with barcode reader YES Operator Signature FB

06300752140000001
Legionella pneumophila 1-6 IgG
CONTROL+ 0.425 mL IVD
LOT 007
2024-11 2024-11 2°C-8°C
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. [Signature]

DATE 05.03.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.