



### QUALITY CONTROL CERTIFICATE

**KIT NAME** tTg-A - CHORUS

Code	86058		
Lot No.	009	2025-02	2024-02

### REAGENTS

Label	Description	Lot.	Exp.	NO
D3431	tTg-A	009	2025-02	6X6
B3431	CALIBRATORE tTg-A	009	2025-02	1
C3431	CONTROL SERUM tTg-A	009	2025-02	1

### CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	370	≥ 100
Control Serum (AU/mL)	39,5	31,7-60,0

### CALIBRATOR

3	3	9	0	0	9	6	2	1	9	3	7	3	8	4	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature FB

339009621937384900  
tTg-A  
CALIBRATOR 0.176 mL IVD  
LOT 009  
2025-02 2°C-8°C  
DIESSE Monteriggioni Italy CE

### CONTROL SERUM

3	3	9	0	0	9	6	2	0	3	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature FB

339009620370000001  
tTg-A  
CONTROL 0.425 mL IVD  
LOT 009  
2025-02 2°C-8°C  
DIESSE Monteriggioni Italy CE

**This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents [Signature]

DATE 07/03/2024

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.