



QUALITY CONTROL CERTIFICATE

KIT NAME Jo-1 - CHORUS

Code	86030		
Lot No.	008	2025-05	2024-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D3261	Jo-1	008	2025-05	6X6
B3261	CALIBRATORE Jo-1	008	2025-05	1
C3261	CONTROL SERUM Jo-1	008	2025-05	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	350	≥ 100
Control Serum (AU/mL)	62,0	28,8-100,0

CALIBRATOR

1	9	7	0	0	8	7	1	2	3	3	4	3	2	4	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*

197008712334324100
Jo-1
CALIBRATOR 0.175 mL [IVD]
LOT 008
2025-05 2°C 8°C
DIESE Monteriggioni Italy CE

CONTROL SERUM

1	9	7	0	0	8	7	2	2	4	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*

197008722410000001
Jo-1
CONTROL 0.425 mL [IVD]
LOT 008
2025-05 2°C 8°C
DIESE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents... *[Signature]*

DATE *08/03/2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.