

**QUALITY CONTROL CERTIFICATE**

**KIT NAME** PR3 - CHORUS

Code	86066		
Lot No.	008	2025-02	2024-02

**REAGENTS**

Label	Description	Lot.	Exp.	NO
D3512	PR3	008	2025-02	6X6
B3512	CALIBRATORE PR3	008	2025-02	1
C3512	CONTROL SERUM PR3	008	2025-02	1

**CONTROLS**

	Found results	Expected results
Calibrator (mOD)	340	≥ 100
Control Serum (AU/mL)	65,4	27,0-100,0

**CALIBRATOR**

2	5	2	0	0	8	6	4	2	5	3	2	3	2	4	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature

252008642532324100  
PR3  
CALIBRATOR 0.175 mL [IVD]  
LOT 008  
2025-02   
DIESSE Monteriggioni Italy

**CONTROL SERUM**

2	5	2	0	0	8	6	2	2	4	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature

252008622410000001  
PR3  
CONTROL+ 0.425 mL [IVD]  
LOT 008  
2025-02   
DIESSE Monteriggioni Italy

**This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents   
DATE 1803.2024

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**  
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.