



QUALITY CONTROL CERTIFICATE

KIT NAME CALPROTECTIN - CHORUS

Code	86600		
Lot No.	008	2025-02	2024-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D9500	CALPROTECTIN	008	2025-02	2 X 6
B9500	CALIBRATORE CALPROTECTIN	008	2025-02	1
C9500	CONTROL CALPROTECTIN	008	2025-02	1
G9502	SAMPLE DILUENT	008	2026-02	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	420	≥ 100
Control (mg/kg)	190,8	100,0-250,0

CALIBRATOR

9	0	0	0	0	8	6	2	1	6	2	1	4	5	2	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

900008621621452900
Calprotectin
CALIBRATOR 0.300 mL IVD
LOT 008
 2025-02 2°C-8°C
DIESE Monteriggioni Italy CE

CONTROL SERUM

9	0	0	0	0	8	6	1	6	2	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

900008616290000001
Calprotectin
CONTROL+ 0.600 mL IVD
LOT 008
 2025-02 2°C-8°C
DIESE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents Anna

DATE 15/03/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.