



QUALITY CONTROL CERTIFICATE

KIT NAME Vitamin D Total- CHORUS

| | | | |
|---------|-------|---------|---------|
| Code | 86900 | | |
| Lot No. | 007 | 2025-02 | 2024-03 |

REAGENTS

| Label | Description | Lot. | Exp. | NO |
|-------|------------------------------------|------|---------|-----|
| D9702 | 25OH Vitamin D Total | 007 | 2025-02 | 6x6 |
| B9703 | CALIBRATOR 25OH Vitamin D Total | 007 | 2026-02 | 1 |
| C9703 | CONTROL SERUM 25OH Vitamin D Total | 007 | 2026-02 | 4 |

CONTROLS


| | Found results | Expected results |
|--|---------------|------------------|
|--|---------------|------------------|

| | | |
|-----------------------|------|-----------|
| Calibrator (mOD) | 565 | ≥ 100 |
| CONTROL SERUM (ng/ml) | 35,4 | 29,4-38,4 |

CALIBRATOR

| | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 8 | 3 | 5 | 0 | 0 | 7 | 6 | 2 | 3 | 2 | 3 | 7 | 5 | 7 | 5 | 6 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|


Label checked with barcode reader YES Operator Signature *FB*



CONTROL SERUM

| | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 8 | 3 | 5 | 0 | 0 | 7 | 6 | 1 | 2 | 2 | 3 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES Operator Signature *FB*



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents: *[Signature]*

DATE 19/03/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.