



DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME BORDETELLA PERTUSSIS TOXIN IgA - CHORUS

Code	81274		
Lot No.	018	2025-03	2024-03

REAGENTS

Label	Description	Lot.	Exp	No
D3440	BORDETELLA PERTUSSIS TOXIN IgA	018	2025-03	6X6
B3440	CALIBRATORE BORDETELLA P.TOXIN IgA	018	2025-06	1
C3440	CONTROL SERUM BORDETELLA P.TOXIN IgA	018	2025-06	1

CONTROLS

	Found results	Expected results
CALIBRATOR (mOD)	512	>100
CONTROL SERUM (IU/ml)	27,0	22,5-40,7

CALIBRATOR

9	6	5	0	1	8	6	2	2	2	3	7	4	4	5	2	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *MB*

965018622237445200
Bordetella pertussis
Toxin IgA
CALIBRATOR 0.480 mL [IVD]
LOT 018
2025-06 2°C-8°C
DIESSE Monteriggioni Italy CE

CONTROL SERUM

9	6	5	0	1	8	6	1	6	3	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *MB*

965018616300000001
Bordetella pertussis
Toxin IgA
CONTROL 0.900 mL [IVD]
LOT 018
2025-06 2°C-8°C
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Bone*

DATE *20.03.2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.