



DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME HERPES SIMPLEX 2 IgG RECOMBINANT - CHORUS

Code	81029		
Lot No.	025	2025-04	2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
DB690	Herpes simplex 2 IgG Recombinant	025	2025-04	6X6
B8690	CALIBRATORE Herpes simplex 2 IgG Recombinant	025	2025-04	1
C8690	CONTROL SERUM Herpes simplex 2 IgG Recombinant	025	2025-04	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	551	≥ 100
Control Serum (INDEX)	1,6	1,2-2,3
Negative Control (INDEX)	0,0	<0,7

CALIBRATOR

2	6	0	0	2	5	6	0	0	0	0	0	5	8	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*



CONTROL SERUM

2	6	0	0	2	5	6	1	2	2	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *[Signature]*
DATE *27/04/2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.