



DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME MEASLES IgM CHORUS

Code	81073		
Lot No.	016	2025-06	2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D2442	MEASLES IgM	016	2025-06	6x6
B2442	CALIBRATORE MEASLES IgM	016	2025-06	2
C2442	CONTROL SERUM MEASLES IgM	016	2025-06	2

CONTROLS

	Found results	Expected results
Calibrator (mOD)	471	≥ 100
Control Serum (Index)	1,9	1,4-2,6
Negative Control (Index)	0,0	< 0,7

CALIBRATOR

1	3	3	0	1	6	7	0	0	0	0	0	5	1	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*

133016700000510000
Measles IgM
CALIBRATOR 0,176 mL [IVD]
LOT 016
2025-06
DIESSE Monteriggioni Italy CE

CONTROL SERUM

1	3	3	0	1	6	7	1	4	2	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*

133016714260000001
Measles IgM
CONTROL 0,425 mL [IVD]
LOT 016
2025-06
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *[Signature]*

DATE *22/04/2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.