

QUALITY CONTROL CERTIFICATE

KIT NAME VARICELLA IgG - CHORUS

Code 81078

Lot No. 025



2025-04



2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D2471	VARICELLA IgG	025	2025-04	6x6
B2472	CALIBRATORE VARICELLA IgG	025	2025-04	1
C2471	CONTROL SERUM VARICELLA IgG	025	2025-04	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	310	≥ 100
Control Serum (Index)	3,8	3,1-5,7
Negative Control (Index)	0,0	$< 0,7$


CALIBRATOR

0	1	4	0	2	5	6	0	0	0	0	0	2	3	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *FB*



CONTROL SERUM

0	1	4	0	2	5	6	3	1	5	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *FB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 23/04/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.