



DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME CHLAMYDOPHILA PNEUMONIAE IgG - CHORUS

Code	81250		
Lot No.	024	2025-04	2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D8590	Chlamydomphila pneumoniae IgG	024	2025-04	6X6
B8590	CALIBRATORE Chlamydomphila p. IgG	024	2025-04	1
C8590	CONTROL SERUM Chlamydomphila p. IgG	024	2025-04	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	470	≥ 100
Control Serum (AU/mL)	41,9	29,1-65,7
Negative Control (AU/mL)	<10,0	≤ 10,0

CALIBRATOR

2	2	2	0	2	4	6	5	1	3	2	4	4	0	4	5	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature MB

222024651324404500
CHLAMYDOPHILA
PNEUMONIAE IgG
CALIBRATOR | 0.175 mL IVD
LOT 024
 2025-04 2°C-8°C
DIESSE Monteriggioni Italy CE 0123
B8590

CONTROL SERUM

2	2	2	0	2	4	6	1	8	3	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature MB

222024618330000001
CHLAMYDOPHILA
PNEUMONIAE IgG
CONTROL+ | 0.425 mL IVD
LOT 024
 2025-04 2°C-8°C
DIESSE Monteriggioni Italy CE 0123
C8590

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents [Signature]

DATE 22/04/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.