DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME CYTOMEGALOVIRUS IgG AVIDITY CHORUS

Code 81092

Lot No. 024 2025-04 2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D2491	CYTOMEGALOVIRUS IgG AVIDITY	024	2025-04	2x6
C2491	CONTROL SERUM HIGH CMV AVIDITY	024	2025-04	1
G2491	SAMPLE DILUENT	024	2025-10	1

CONTROLS

	Found results	Expected results
Control Serum (%)	69,7%	50-95%

STICK LABEL

CALIBRATOR

Label checked with barcode reader YES ☐ Operator Signature .....

CONTROL SERUM

0 8 4 0 2 4 6 5 0 9 5 0 0 0 0 0 0 1

Label checked with barcode reader YES ☒ Operator Signature .....

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 23-06-2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma.All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.